F.T.R. RESEARCH AND DEVELOPMENT
CH - 2003 NEUCHATEL / SWITZERLAND
INTER - OFFICE CORRESPONDENCE

To: See distribution Date: 18.1.1995

From: S. Renaud Doc name: ct1294.doc

Distribution:	A. Mengoni	PME-Neuchâtel
	B. Strang	PM-USA
	J. De Wilde	PMH-Bergen op Zoom
	K. Wagner	PMG-Munich
Copies:	T. Bel	PME-Neuchâtel
	R. Hofer	PME-Neuchâtel
	D. Parthenopoulos	PME-Neuchâtel
	M. Rigaux	PME-Neuchâtel
	S. Terao	PME-Neuchâtel
	D. Turpin	PME-Neuchâtel

The IM14 results are given on two different data sheets, the first one, **MONTHLY RESULTS : MONITOR**, contains the average results of the IM-14 smoked according to ISO during the month and the average OV values of the MLK-15. The second one, **INTERLABORATORY COLLABORATIVE TEST : IM 14**, contains the average values of the monitors smoked during the collaborative test.
The results obtained for the chemical analysis of the Monitor Tobacco are listed in two tables, the first one, **MONTHLY RESULTS : MONITOR TOBACCO**, contains the results obtained during the month, while the second one, **INTERLABORATORY COLLABORATIVE TEST : MONITOR TOBACCO**, lists the results obtained during the collaborative test.

GENERAL COMMENTS

Due to technical problems PME did not report any results of the chemical analysis for the collaborative test in December.
R&D's TA and RS results of the monthly analysis of the Monitor Tobacco for September 1994 were corrected.
In November, the smoke deliveries for the MLB increased and in December they decreased again. All labs detected these changes closely.
Since November all labs detect an increase in RS for the MER.

Source: https://www.industrydocuments.ucsf.edu/docs/lzlk0001

RESULTS OF THE LABORATORIES

PME

PME reported high smoking yields for the IM-14 smoked on Filtrona during the collaborative test in November and low ones in December. On Borgwaldt, PME reported lower PC's than PMG for all cigarettes analyzed. The difference in water results between PMG and PME on Borgwaldt persists for the monitor and all brands with the exception of the MLY in November.
RS results are low in November for all brands.
In December PME reported the lowest Filter RTD's for all brands and the highest ventilation results, except for the MLB.

PMH

PMH reported low TPM and tar results for the IM-14 smoked during the collaborative test in November. PMH's water and OV results are somewhat lower for all brands, except for the MLY, in November and December, which results in lowered PC's.
RS results are high for the Monitor Tobacco and all brands in December, except for the MLB.

PMG

PMG is on the high side for water results on Filtrona in November and for TPM and tar values in December.
$N\text{-}NO_3$ results are on the high side for all brands and the Monitor Tobacco both months. $N\text{-}NH_3$ results are somewhat high for all brands in December. For the Monitor Tobacco the RS content is still low.

R&D

R&D's OV results are again high in November and December. TPM, tar and SN results are also somewhat high in November for the brands and the IM14 smoked during the test. This could again be correlated with the higher PC's.
In November, R&D is high for TA reported for all brands with the exception of the PMU. In December the same is true with the exception of the MER. R&D is low for the ventilation and the cigarette weight in November.

S. Renaud

Sabine Renaud

Source: https://www.industrydocuments.ucsf.edu/docs/lzlk0001